



Product Service

# Attestation of Conformity

No. D4 10 10 40489 010

**Holder of Certificate:** Hwashin Technology Co.216-7, Jangji-dong  
Gwangju-si, Gyeonggi-do 464-070  
REPUBLIC OF KOREA**Product:** Ultrasonic cleaners  
(Ultrasonic cleaners)**Model(s):** POWERSONIC 405, POWERSONIC 410, POWERSONIC 420,  
POWERSONIC 505, POWERSONIC 510, POWERSONIC 520,  
POWERSONIC 603  
(Differences : Cleaner Tank Capacity, Power Consumption,  
Control Panel, Size, Enclosure Color and Weights)**Parameters:** Rated Power : 230Va.c., 50Hz  
Power Consumption : 350W/500W/700W  
LCD Display : POWERSONIC 500 Series  
Oscillator : BLT Type  
Oscillating Frequency : 40kHz  
Time Setting : Max. 99 Minutes  
Bath Capacity(Liters) : 3.3 / 5.7 / 10.0 / 20.7**Tested according to:** IEC 62321:2008

Pb, Hg, Cd, Cr(VI), PBBs and PBDEs could not be detected over the limit defined by the European Directive 2002/95/EC.

This Attestation of Conformity confirms the compliance with the listed standards on a voluntary basis. It refers only to the sample submitted and does not certify the quality or safety of the serial products. See also notes overleaf.

**Test report no.:** CPSD0330340**Date,** 2010-10-11

( James Jeon )



Page 1 of 1